



## PRODUCT SPECIFICATIONS

Collagenase MA, Catalog #001-2030

Version 16 Dec 2016

**Product:** Aseptically dispensed mixture of purified Class 1 (60%) and Class 2 (40%) Collagenase from *C. histolyticum*

Lot #	SPECIFICATION	RESULT
Appearance	White lyophilized cake	Lot Specific
Specific Collagen Degrading Activity <sup>1</sup>	15,000 – 30,000 Units/mg	Lot Specific
Total Collagen Degrading Activity <sup>1</sup>	2.25 – 2.75 Million Units/bottle	Lot Specific
Total Protein <sup>2</sup>	Report Only	Lot Specific
Endotoxin <sup>3</sup>	< 10.0 EU/mg	Lot Specific

**STORAGE:** < -20°C

**DATE OF MANUFACTURE:**

**EXPIRY DATE:** 2 years from date of manufacture

<sup>1</sup> McCarthy RC, et. al. Development and Characterization of a Collagen Degradation Assay to Assess Purified Collagenase Used in Islet Isolation. *Transplantation Proceedings* 40 (2008); 339-342

<sup>2</sup> Based on absorbance at 280 nm using an extinction coefficient of  $\epsilon=1.41$

<sup>3</sup> Charles River Endosafe® PTS Endotoxin Assay

**VitaCyte, LLC**

1341 Sadlier Circle West Drive

Indianapolis, IN 46239

www.vitacyte.com

**Phone:** 317.917.3457

**Fax:** 317.917.3459